

091205T4PTT

PERIOPERATIVE THEATRE TECHNOLOGY 5

HE/OS/TT/CR/04/5/A

Provide Perioperative Theatre Technical Services

Nov/Dec 2023



**TVET CURRICULUM DEVELOPMENT, ASSESSMENT AND CERTIFICATION
COUNCIL (TVET CDACC)**

WRITTEN ASSESSMENT

Time: 3 Hours

INSTRUCTIONS TO CANDIDATES

1. This paper has three sections **A, B** and **C**.
2. You are provided with a separate answer booklet.
3. Marks for each question are as indicated.
4. Do not write on the question paper.

This paper consists of 7 printed pages.

Candidates should check the question paper to ascertain that all pages are printed as indicated and that no questions are missing.

SECTION A: (20 MARKS)

*Answer **all** questions in this section.*

Each question carries one Mark.

1. The absence of all forms of microbial life, including spores, is known as
 - A. Sensitization
 - B. Disinfection
 - C. Decontamination
 - D. Sterility
2. Where are test packs placed to present the greatest challenge to the steam sterilizer.
 - A. Bottom rack back
 - B. Top rack back
 - C. Center
 - D. Over the drain
3. With respect to principles of surgical asepsis, when you are in doubt about the sterility of an item, you should:
 - A. “Throw it out”
 - B. “Test its sterility”
 - C. “Ignore the red flags”
 - D. “Sterilize it”
4. Which of the following is not a suitable type of eye protection in healthcare?
 - A. Prescription glasses
 - B. Polycarbonate protective glasses
 - C. Goggles
 - D. Face shields
5. When must you remove and change out of single use PPE?
 - A. After completing a ward round
 - B. After any patient contact
 - C. At the end of your shift
 - D. Only if it gets damaged
6. How can medical gloves be made reusable?
 - A. With an alcohol-based hand rub
 - B. They can't - medical gloves must always be disposed of after use
 - C. Washing them with water and soap