

091205T4PTT

PERIOPERATIVE THEATRE TECHNOLOGY 5

HE/OS/TT/CR/03/5/A

Maintain Operation Theatre Instruments and Apparatus

Nov/Dec 2023



**TVET CURRICULUM DEVELOPMENT, ASSESSMENT AND CERTIFICATION
COUNCIL (TVET CDACC)**

INSTRUCTIONS TO THE ASSESSOR

1. You are required to mark the practical as the candidate perform the tasks.
2. You are required to take video clips at critical points.
3. Ensure the candidate has a name-tag and registration code at the back and front.

This paper consist of 3 printed pages.

Assessor should check the question paper to ascertain that all pages are printed as indicated and that no questions are missing.

OBSERVATION CHECKLIST

Candidate's name & Registration No.			
Assessor's name & Reg. code			
Unit(s) of Competency			
Venue of Assessment			
Date of assessment			
Items to be evaluated: <i>(Award marks appropriately as guided for in the items for evaluation indicated. Give a brief comment where necessary)</i>	Marks allocated	Marks obtained	Comments
1. Washed hands as per WHO guidelines	1		
2. Donned appropriate personal protective equipment a) Donned a pair of cleaned gloves b) Put on a plastic apron c) Wore theatre boots d) Put on face mask <i>(Award 4 marks for correct use of all the above PPE or 0)</i>	4		
3. Assembled instruments from the tables a) Checked drapes, towels and table covers b) Collected instruments from the mayo table and other tables c) Confirmed that there was no dropped items and collected if there were any that had dropped <i>(Award 1 marks for each)</i>	3		
4. Sorted medical apparatus a) Separated delicate instruments, small items and sharps from other instruments b) Disassembled all instruments with removable parts c) Separated instruments of dissimilar metals <i>(Award 1 marks for each)</i>	3		

5. Put the sorted equipment in the prepared disinfectant <i>(Award 1 mark or zero)</i>	1		
6. Allowed time (15 minutes) for instruments contact with the disinfectant as per the manufacturers policy <i>(Award 1 mark or 0)</i>	1		
7. Flushed cold water through hollow instruments <i>(Award 1 mark or 0)</i>	1		
8. Rinsed off debris and blood with water from the non-hollow equipment <i>(Award 1 mark or 0)</i>	1		
9. Dipped and scrubbed instruments in soapy water <i>(Award 1 mark or 0)</i>	1		
10. Rinsed instrument in clean water	1		
11. Prepared apparatus for sterilization a) Set the instruments in a drying area and wait for them to dry b) Wrapped the instruments in readiness for sterilization <i>(Award marks as indicated for each)</i>	1 2		
Total	20		
The candidate was found to be: Competent <input type="checkbox"/> Not yet competent <input type="checkbox"/> <i>(Please tick appropriately)</i> <i>(The candidate is competent if s/he gets 50% and above.)</i>			
Feedback from candidate:			
Feedback to candidate:			
Candidate's signature:	Date:		
Assessor's signature:	Date:		

